

TEST REPORT

by accredited test centre
TÜV No.: M/MG3-W/15-001
Unit Verification Test

TÜV AUSTRIA
SERVICES
GMBH

Branch office:
Deutschstrasse 10
1230 Vienna
Phone:
+43(0)1 610 91-()
Fax: DW 6505
pzw@tuv.at

Business Area:
Medical technology/
Communications
technology/ EMC
section

Contact person:
Dipl.-Ing.
Andreas
PARZER
DW6510
par@tuv.at

TÜV^o

Requester: **HEALTHFACTORIES GmbH**
D-83395 Freilassing, Fürstenweg 4

Test item: **Magnetic field therapy device for horses**

Mark of origin: **HEALTHFACTORIES GmbH**
D-83395 Freilassing, Fürstenweg 4

Type: **EQUITRON** Serial number: **15 1000 0001**

Supply voltage: **230V~** Supply frequency: **50 Hz**

Power input/consumption: **170VA** Operating mode: **Continuous operation**

Protection class: **I** Degree of protection: **BF, conventional**

Test based on: **EN 60601-1: 2006 + Corr. 1 + Corr.2 + A1:2013**

This test report only applies to the individual device listed above with the serial number stated.

The above medical electrical device (single item) was subjected to a unit verification test of compliance with the above test principles by TÜV Austria Services GmbH. No potentially destructive, long-term or heat tests or tests for electromagnetic compatibility were performed so that this test does not represent a complete type test.

The process of risk management was not assessed in detail. Only those aspects of risk management which are connected with the examinations were tested. There is an overview of the tests performed on page 2 of the report. Test reports are available from the test centre.

ASSESSMENT

No safety or functional defects were found.

(Any instructions listed below must be complied with during operation).

The device was labelled with a "Tested by TÜV AUSTRIA SERVICES GMBH" sticker.

TÜV AUSTRIA SERVICES GMBH
Medical technology / Communications technology / EMC

Date	Responsible person	Authorised Signatory:
27/01/2016	[Signature] Dipl.-Ing. Andreas Parzer	[Signature] Dipl.-Ing. Michael Pölzleitner

Accredited as:

Test centre,
inspection centre,
calibration centre,
verification centre,
initial and boiler
test centre

Notified Body 0408

**Chair of the
supervisory board:**
KR Dipl.-Ing. Johann
MARIHART

Directors:
Dipl.-Ing. Dr.
Stefan HAAS
Mag. Christoph
WENNINGER

Head office:
Krugerstrasse 16
1015 Vienna/Austria

**Further branch
offices** Dornbim, Graz,
Innsbruck, Klagenfurt,
Linz, Salzburg, St..
Pölsen, Wels, Vienna,
Brixen (I) and Filderstadt

**Commercial register
court number:**
Vienna / FN 288476 f

Bank details:
UC BA 52949 001 066
IBAN
AT131200052949001066
BIC BKAUATWW
RZB 001-04.093.282
IBAN
AT153100000104093282
BIC RZBAATWW

UID ATU63240488
DVR3002476